



बी.पी. कोइराला स्वास्थ्य बिज्ञान प्रतिष्ठान, धरानको

कोटेशन माग गरिएको सुचना

सुचना प्रकाशित मिति: २०८०/१२/०२

यस प्रतिष्ठानलाई आवश्यक तपशिलका उल्लेखित **Centrifuge Machine** खरिद गर्नुपर्ने भएकोले इजाजत प्राप्त प्रतिष्ठानमा सुचिकृत फर्म, संस्था, कम्पनीबाट अद्यावधिक फर्म दर्ता प्रमाणपत्र, मु.अ.कर दर्ता प्रमाणपत्र, आ.व.२०७९/०८० को कर चुक्ताको प्रमाणपत्र प्रतिहरु संलग्न गरी यो सुचना प्रकाशित मितिले ६ दिन भित्र कोटेशन पेश गर्नुहुन सुचित गरिन्छ। माग गरिएको कोटेशन तोकिएको म्यादभित्र प्रतिष्ठानको दर्ता चलानी फाँटमा दर्ता गर्नुहुन वा quotation.procurement@bpkihs.edu मा email मार्फत पेश गर्नुपर्नेछ।

SN.	Product Name	Quantity	Unit
1	Centrifuge Machine	2	Nos

Laboratory centrifuge with rotor (Two units)

Description: General Purpose portable Laboratory Centrifuge with swing out rotor and brushless induction motor.

Specifications:-

- 1) Rotor capacity 32 x 15 ml swing out.
- 2) Maximum speed should be at least 6000rpm/RCF-4390 g .
- 3) Power supply 230 volt +/-10v, 50 Hz AC
- 4) Should have automatic rotor identification
- 5) Should have stainless steel centrifuge chamber
- 6) Brushless induction motor with variable frequency drive
- 7) Microprocessor controller with digital display
- 8) Programmable speed regulator or Stable speed output even under unstable voltage condition
- 9) Smooth and soft start, low temperature rise
- 10) LED display of speed
- 11) Digital countdown timer and continuous run (range 0-99 mins)
- 12) Safety lid interlock to prevent lids opening during centrifugation/electronic safety lid lock
- 13) Imbalance detector and centrifugation stop with display of error
- 14) Dynamic break for quick declaration
- 15) Motor over head protection
- 16) Gas hinge to prevent door falling
- 17) Emergency lid lock release
- 18) Heavy Duty C.R.C. Construction finished with power coating
- 19) Must be three years warranty
- 20) Must provide spare parts and service after warranty.
- 21) Must have US FDA approval or CE certified.
- 22) System Noise must be Less than 65 dB
- 23) Country origin should be mentioned.