



वी.पी.कोइराला स्वास्थ्य विज्ञान प्रतिष्ठान धरानको
कोटेशन माग गरिएको सूचना ।

(सूचना प्रकाशित मिति: २०८१।१२।०८)

यस प्रतिष्ठानको लागि आवश्यक संलग्न Specification बमोजिमको Surgical Headlight- ३ थान खरिद गर्न लागत अनुमान तयार गर्नुपर्ने भएकोले सो प्रयोजनको लागि व्यवसाय (फर्म) दर्ता प्रमाणपत्र, मु.अ.कर दर्ता प्रमाणपत्र, आ.व. २०८०/०८१ को कर चुक्ताको प्रमाणपत्रको प्रतिहरू संलग्न गरी यो सूचना प्रकाशित मितिले ७ (सात) दिन भित्र कोटेशन तथा क्याटलगहरू पेश गर्नुहुन सुचित गरिन्छ । माग गरिएको कोटेशन तथा क्याटलगहरू तोकिएको म्यादभित्र प्रतिष्ठानको दर्ता, चलानी फाँटमा दर्ता गर्नुहुन वा quotation.procurement@bpkihs.edu मा email मार्फत पेश गर्नुपर्नेछ ।

1. Surgical Head Light

Technical Specifications

1. Light output: 140,000 Lux or above at 16 inches
2. Spot size: freedom of spot size adjustability and control or a minimum of 66.5mm @ 16 inches.
3. Weight: should be light weight. (80g – 310g)* preferable on the lower range.
4. Color Temperature: above 5000K with accurate color reproducibility.
5. Should have greater brightness consistency to reduce light fall – off or should have high definition imaging (HDI) for uniform light distribution.
6. Should have optically collimated two LED light source mounted on a head band.
7. Should have additional batteries provided within the kit.
8. Batteries should be rechargeable.
9. Head band should be light and should have cushion within for optimal comfort.
10. Should be cordless.
11. Certification: Should be USFDA and European CE certified.