

## वी.पी. कोइराला स्वास्थ्य बिज्ञान प्रतिष्ठान, धरानको

### कोटेशन माग गरिएको सुचना

(सुचना प्रकाशित मिति २०८०/०४/२२)

(पुनःसुचना प्रकाशित मिति २०८०/०५/०४)

यस प्रतिष्ठानको अस्पताल फार्मेशीमा बिक्री बितरण प्रयोजनको लागि आवश्यक **Chest Drain System 100pcs** कोटेशनको माध्यमबाट खरिद गर्नुपर्ने भएकोले इजाजत प्राप्त प्रतिष्ठानमा सुचिकृत फर्म, संस्था, कम्पनीबाट अद्यावधिक फर्म दर्ता प्रमाणपत्र, मु.अ.कर दर्ता प्रमाणपत्र, आ.व.२०७९/०८० को कर चुक्ताको प्रमाणपत्र प्रतिहरु संलग्न गरी यो सुचना प्रकाशित मितिले ७ दिन भित्र कोटेशन पेश गर्नुहुन सुचित गरिन्छ । माग गरिएको कोटेशन तोकिएको म्यादभित्र प्रतिष्ठानको दर्ता चलानी फाँटमा दर्ता गर्नुहुन वा [quotation.procurement@bpkihs.edu](mailto:quotation.procurement@bpkihs.edu) मा email मार्फत पेश गर्नुपर्नेछ ।

### Specification for chest drainage system

1. Sterile underwater seal chest drainage system.
2. Inlet tube- should be translucent, flexible and able to milk and should be able to fit all sizes of chest tubes.
3. Collection chamber should of two types- one chamber system for connection of one chest tube and two chamber system for connection of two tubes.
4. Capacity of collection chamber should be > 2litres.
5. Should be able to connect in all vacuum sources at different negative pressures without any water migration from chambers.
6. Low pressure negative suction regulator to regulate the applied negative pressure from wall vacuum source. The regulator setting should vary from - 5cm H<sub>2</sub>O to -40cm H<sub>2</sub>O pressure.
7. Should have self sealing sampling port for sample collection